



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Metrological COM IN TEC Services, S.C.***  
*Calle Zacamixtle # 108 Col. Petrolera*  
*Delegación Azcapotzalco, Ciudad de México, México. C.P. 02480*

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Optical, Chemical, Dimensional, Thermodynamic, Mass, Force and Weighing  
Devices and Mechanical Calibration***  
*(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

*Initial Accreditation Date:*

*Issue Date:*

*Expiration Date:*

July 02, 2013

September 12, 2021

November 30, 2023

*Accreditation No.:*

*Certificate No.:*

71793

L21-555

Tracy Szerszen  
President

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjlab.com](http://www.pjlab.com)*



# Certificate of Accreditation: Supplement

## Metrological COM IN TEC Services, S.C.

Calle Zacamixtle # 108 Col. Petrolera  
 Delegación Azcapotzalco, Ciudad de México, México C.P. 02480  
 Contact Name: María del Refugio Castañeda Avelar Phone: 555-369-4971

Accreditation is granted to the facility to perform the following calibrations:

### Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Reflectance Color <sup>FO</sup> Spectrometers, Reflectance Geometric d/0°	CIE L*: 0 to 100 CIE a*: -60 to 60 CIE b*: -60 to 60	L*: 0.36 a*: 0.31 b*: 0.12	Ceramic Tiles Konica Minolta Model: BCRA CENAM Technical Guide
Reflectance Color <sup>FO</sup> Spectrometers, Geometric 45/0° CIE Lab	400 nm to 700 nm 0 % reflectance to 100 % reflectance  CIE L*: 0 to 100 CIE a*: -80 to 80 CIE b*: -80 to 80	1.2 % reflectance  L*:0.11 a*: 0.08 b*: 0.06	
Reflectance Color <sup>FO</sup> Spectrometers, Geometric d/8 CIE Lab	400 nm to 700 nm 0 % reflectance to 100 % reflectance  CIE L*: 0 to 100 CIE a*: -80 to 80 CIE b*: -80 to 80	0.9 % reflectance  L*: 0.22 a*: 0.15 b*: 0.04	
Ceramic Color in space CIE Optical Geometry <sup>F</sup> d/8°	0 % reflectance to 100 % reflectance  CIE L*: 0 - 100 CIE a* -75 to 75 CIE b* -75 to 75	1 % reflectance  L* 0.23 a* 0.16 b* 0.05	Spectrophotometer Konica Minolta with Optical Geometry d/8° ASTM C609-20
Ceramic Color in Space <sup>F</sup> Optical Geometry 45°/0°	0 % RH to 100 % RH  CIE L*: 0 - 100 CIE a* -75 to 75 CIE b* -75 to 75	1.3 % reflectance  L* 0.12 a* 0.09 b* 0.06	Spectrophotometer Konica Minolta with Optical Geometry 45°/0° ASTM C609-20
Ceramic Color in Space CIE Optical Geometry <sup>F</sup> d/0°	CIE L*: 0 - 100 CIE a* -75 to 75 CIE b* -75 to 75	L* 0.51 a* 0.41 b* 0.26	Spectrophotometer Konica Minolta with Optical Geometry d/0° ASTM C609-20
Transmittance Spectrophotometers <sup>FO</sup>	10 % T to 50 % T Spectral Bandwidth (1 n-m)	0.036 % T	Neutral Density Glass Filters, Interference Filters CENAM Technical Guide
Gloss Meters <sup>FO</sup> Fixed Points	Angle 20°: 94 Gloss Units Angle 60°: 96 Gloss Units Angle 85°: 100 Gloss Units	0.17 Gloss Units 0.19 Gloss Units 0.2 Gloss Units	High Gloss Glass ASTM D-523-14 (2018)



# Certificate of Accreditation: Supplement

## Metrological COM IN TEC Services, S.C.

Calle Zacamixtle #. 108 Col. Petrolera  
 Delegación Azcapotzalco, Ciudad de México, México C.P. 02480  
 Contact Name: María del Refugio Castañeda Avelar Phone: 555-369-4971

Accreditation is granted to the facility to perform the following calibrations:

### Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Gloss Ceramic Tile <sup>F</sup>	Angle 20°: Up to 2 000 Gloss Units	0.15 Gloss Units	Gloss Meter Konica Minolta ASTM D-523-14 (2018)
	Angle 60°: Up to 1 000 Gloss Units	0.17 Gloss Units	
	Angle 85°: Up to 160 Gloss Units	0.19 Gloss Units	
Ev Illuminance <sup>FO</sup>	10 Lux to 2 900 Lux	1.3 Lux	Light Meter
Ev Light Color <sup>FO</sup>	2 856 K	5.8 K	Konica Minolta CL-200A ASTM D1729-16

### Chemical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Dynamic Viscometers Rotational <sup>FO</sup>	1 000 mPa·s	4 mPa·s	Viscosity Standards Cannon Technical Guide CENAM
	5 000 mPa·s	21 mPa·s	
	12 500 mPa·s	55 mPa·s	
pH Meters (Potential of Hydrogen) <sup>FO</sup>	4 pH to 10 pH	0.012 pH	pH Buffer Solutions Technical Guide CENAM
Conductivity Meters Fixed Points <sup>FO</sup>	100 $\mu$ S/cm	0.4 $\mu$ S/cm	Conductivity Solutions CENAM Technical Guide
	1 408 $\mu$ S/cm	3.3 $\mu$ S/cm	
Kinematic Viscosity <sup>F</sup>	118.5 mm <sup>2</sup> /sec	0.34 mm <sup>2</sup> /sec	Viscosity Standards Cannon CENAM Technical Guide
	396.5 mm <sup>2</sup> /sec	1.2 mm <sup>2</sup> /sec	
CAP Type Viscometer Calibration / Rheometer <sup>F</sup>	3.1 mPa·s to 29.78 mPa·s	0.3 % of reading	Paragon Viscosity Standards ASTM D4287-00
	6.32 mPa·s to 138.8 mPa·s	0.31 % of reading	
	385.3 mPa·s to 551.2 mPa·s	0.32 % of reading	Cannon Viscosity Standards ASTM D4287-00
	759.6 mPa·s to 1 083 mPa·s	0.33 % of reading	



# Certificate of Accreditation: Supplement

## Metrological COM IN TEC Services, S.C.

Calle Zacamixtle #. 108 Col. Petrolera  
 Delegación Azcapotzalco, Ciudad de México, México C.P. 02480  
 Contact Name: María del Refugio Castañeda Avelar Phone: 555-369-4971

Accreditation is granted to the facility to perform the following calibrations:

### Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Thickness Gages <sup>FO</sup>	49 $\mu$ m	0.41 $\mu$ m	Thickness Gages CENAM Technical Guide
	351 $\mu$ m	0.41 $\mu$ m	
	977 $\mu$ m	0.41 $\mu$ m	
Micrometers <sup>F</sup>	0.5 mm to 252 mm	0.001 3 mm	Master Gage Blocks
Calipers <sup>F</sup>	0.5 mm to 252 mm	0.01 mm	CENAM Technical Guide

### Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Liquid in Glass Thermometer <sup>F</sup> (Partial Immersion)	25 °C to 100 °C	0.7 °C	RTD Digital and Temperature Bath CENAM Technical Guide
	100 °C to 150 °C	0.71 °C	
Bimetal Thermometer <sup>F</sup>	25 °C to 100 °C	0.76 °C	
	100 °C to 200 °C	0.77 °C	
Indicators Temperature with Thermocouple Type E <sup>FO</sup>	25 °C to 100 °C	0.54 °C	
	100 °C to 200 °C	0.54 °C	
	200 °C to 300 °C	0.57 °C	
Indicators Temperature with Thermocouple Type J <sup>FO</sup>	25 °C to 100 °C	0.52 °C	
	100 °C to 200 °C	0.52 °C	
	200 °C to 300 °C	0.55 °C	
Indicators Temperature with Thermocouple Type K <sup>FO</sup>	25 °C to 100 °C	0.53 °C	
	100 °C to 200 °C	0.53 °C	
	200 °C to 300 °C	0.54 °C	
Indicators Temperature with Thermocouple Type T <sup>FO</sup>	25 °C to 100 °C	0.53 °C	RTD Digital and Temperature Bath CENAM Technical Guide
	100 °C to 200 °C	0.53 °C	
	200 °C to 300 °C	0.55 °C	
Digital Thermometer <sup>FO</sup>	5 °C to 400 °C	0.48 °C	RTD Digital and Dry Well CENAM Technical Guide
Termohygrometer Temperature <sup>F</sup>	5 °C to 60 °C	0.26 °C	RTD Digital and Chamber Climatic CENAM Technical Guide



# Certificate of Accreditation: Supplement

## Metrological COM IN TEC Services, S.C.

Calle Zacamixtle #. 108 Col. Petrolera  
 Delegación Azcapotzalco, Ciudad de México, México C.P. 02480  
 Contact Name: María del Refugio Castañeda Avelar Phone: 555-369-4971

Accreditation is granted to the facility to perform the following calibrations:

### Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Termohygrometer Humidity <sup>F</sup>	10 % HR to 80 % HR	0.78 % HR	Hygrometer Digital CENAM Technical Guide

### Mass, Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Analytical Balances <sup>O</sup>	1 mg to 200 g (Res.= 0.1 mg)	0.3 mg	Class OIML E2 Weights CENAM Technical Guide
Balances <sup>O</sup>	10 mg to 500 g (Res.= 0.2 mg)	0.7 mg	
	200 g to 5 000 g (Res.= 0.005 g)	6.3 mg	Class OIML M1 Weights CENAM Technical Guide
	5 kg to 10 kg (Res.= 0.1 g)	0.6 g	
Scales <sup>O</sup>	10 kg to 100 kg (Res.= 20 g)	18 g	
	100 kg to 200 kg (Res.= 20 g)	18 g	
	100 kg to 200 kg (Res.= 10 g)	10 g	
	100 kg to 250 kg (Res.= 20 g)	18 g	
	200 kg to 300 kg (Res.= 50 g)	42 g	
Mass Weight Class F1, F2 <sup>F</sup>	1 g	0.007 mg	Double Substitution Class E2 Weights Set CENAM Technical Guide
	2 g	0.015 mg	
	5 g	0.018 mg	
	10 g	0.021 mg	
	20 g	0.028 mg	
	50 g	0.034 mg	
	100 g	0.078 mg	
	200 g	0.12 mg	
	500 g	0.64 mg	
	1 kg	0.79 mg	
2 kg	1.2 mg		
5 kg	6.4 mg		



# Certificate of Accreditation: Supplement

## Metrological COM IN TEC Services, S.C.

Calle Zacamixtle #. 108 Col. Petrolera

Delegación Azcapotzalco, Ciudad de México, México C.P. 02480

Contact Name: María del Refugio Castañeda Avelar Phone: 555-369-4971

Accreditation is granted to the facility to perform the following calibrations:

### Mass, Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Mass Weight Class M1, M2, M3 <sup>F</sup>	5 kg	6.4 mg	Double Substitution Class F2 Weights Set CENAM Technical Guide
	10 kg	79 mg	
	20 kg	116 mg	
Force - Tension Instruments (Dynamometer, Universal Machine and Load Cells) <sup>FO</sup>	20 N to 1 000 N	$(3.93 \times 10^{-3} + 7.75 \times 10^{-3}F)$ N	OIML Class M1 ASTM E4 CENAM Technical Guide
Force – Compression Instruments (Dynamometer, Universal Machine and Load Cells) <sup>FO</sup>	20 N to 1 000 N	$(3.93 \times 10^{-3} + 7.75 \times 10^{-3}F)$ N	

### Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Vacuum Gauges <sup>FO</sup>	-12 psi to 0 psi	0.35 psi	Digital Pressure Gauge CENAM Technical Guide
Pressure Gauges and Transducer <sup>FO</sup>	Up to 300 psi	0.021 psi	

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor  $k$  (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this calibration at its fixed location.
4. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations. Example: Outside Micrometer<sup>O</sup> would mean that the laboratory performs this calibration onsite at the customer's location.



## Certificate of Accreditation: Supplement

### Metrological COM IN TEC Services, S.C.

Calle Zacamixtle #. 108 Col. Petrolera

Delegación Azcapotzalco, Ciudad de México, México C.P. 02480

Contact Name: María del Refugio Castañeda Avelar Phone: 555-369-4971

*Accreditation is granted to the facility to perform the following calibrations:*

5. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer<sup>FO</sup> would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
6. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
7. The term T represents torque in N•m (including SI multiple and submultiple units) for the international system of units (the SI) or ozf•in, lbf•in and lbf•ft for the USC system of units.
8. The term F represents Force in N (including SI multiple and submultiple units) for the international system of units (the SI) or lbf for the USC system of units.
9. This is the parent location.

